

Ensuring high quality ABR in babies

Guy Lightfoot

English NHSP Clinical Group

email: guy.lightfoot@nhs.net



NHSP Mission Statement

'Improving outcomes for every child through a high quality hearing screening programme, safe and effective assessments and family centred intervention'

- Programme roll-out 2001 – 2006
- Statistics: in a typical week
 - 13,200 babies are screened
 - 296 (2%) are referred for audiological assessment
 - 22 babies are identified with a permanent childhood hearing impairment (PCHI)


Infrastructure

- Screening and assessment protocols
 - TEOAE initial screen (AABR+TEOAE for high risk)
 - AABR for screening referrals
 - Diagnostic ABR with 30 dBeHL discharge level
 - Protocols for Tone pip, click AC & BC ABR, CM testing
- Staff training & support
 - Specialist training courses for screeners & ABR staff
- Quality Standards
- Quality Assurance monitoring (audits)

QA audit findings

- 2006/8 audit revealed that the quality of ABRs at some sites was poor and risked missing some babies with PCHI
- Why?
 - Unconscious incompetence of ABR tester
(“I’ve been doing this for years; I don’t need to be told what to do”)
 - Too small a population per tester
(doing tests too infrequently to acquire & maintain competence)
 - Failure by tester to follow national ABR test protocols
(“I’m reasonably sure I can see a response there”)

Threshold ABR Left



Sensitivity and Sweep Time Per Division

1 | 0.25 uV 1.5 msec

2 | 0.25 uV 1.5 msec


3 | 0.25 uV 1.5 msec

4 | 0.25 uV 1.5 msec

5 | 0.25 uV 1.5 msec

6 | 0.25 uV 1.5 msec

Threshold ABR Left






Sensitivity and Sweep Time Per Division

1 | 0.25 uV 1.5 msec 2 | 0.25 uV 1.5 msec 3 | 0.25 uV 1.5 msec 4 | 0.25 uV 1.5 msec 5 | 0.25 uV 1.5 msec 6 | 0.25 uV 1.5 msec

Examples of discharged cases

(35 dBnHL click ABR reported as present)



Outline of current ABR protocol

- Test Parameters

- 20ms timebase (25ms for 1kHz & 500Hz)
- Rate: as fast as the timebase allows (49.1/s; 39.1/s)
- Filters: 30Hz – 1500Hz
- Sweeps: typically 3000 (min 2000)
- Artifact rejection: $\pm 5 - \pm 10 \mu V$
- Display scale: fixed; $0.05 - 0.1 \mu V = 1ms$
- Replication: needed at stimulus levels that define the ABR threshold
- 4kHz primary stimulus, upon which discharge is based

Definition of ABR threshold

The lowest stimulus level at which there is:

- a replicated clear response at threshold
 - a replicated clear response at 10dB or 5dB above threshold, and
 - a replicated recording with response absent at 10dB or 5dB below threshold.
-
- Replicated waveforms should be “optimally superimposed” – shown on the same baseline rather than adjusted to that peaks are aligned

Interpretation of waveforms


“Clear Response” (CR) requirements:

- High degree of correlation between the replications and a characteristic waveform of at least $0.05\mu\text{V}$ (50nV) in size
- The size of the response should be at least *3 times the amplitude of the background noise level*
- The noise level can be estimated from average gap between the traces across the recording window

This criterion ensures a high degree of confidence

(about 98%) in the presence of an ABR response


Example – Clear Response present



Interpretation of waveforms

“Response Absent” (RA) Requirements


- Noise is the average gap between replicates (ignore any region of stimulus artefact).
- Average gap must be no more than 25nV (0.025 μ V).



Interpretation of waveforms


All other waveforms are “inconclusive” (Inc)

- the replications will have $S/N < 3:1$ or have no obvious response yet have noise greater than 25nV




Inconclusive waveforms cannot contribute to the definition of the ABR threshold!

The result: high quality, fewer errors




Discharge level reached at 4kHz bilaterally

The result: high quality, fewer errors




Discharge level reached at 4kHz bilaterally

The result: high quality, fewer errors




Discharge level reached at 4kHz & 1kHz

The result: high quality, fewer errors



A mild unilateral cochlear loss (BC was masked)

The result: high quality, fewer errors



A profound bilateral loss (very low residual noise)

The most serious errors in practice

1. Waiting until the baby is too old
Plan to complete testing by 8 weeks
2. Amp sensitivity too low (artifact rejection too lax)
NHSP advice: $\pm 5\mu\text{V}$ (max $\pm 10\mu\text{V}$)
3. Using an incorrect display scale
4. Failure to replicate at levels that define threshold
Test at threshold, at 5-10dB below AND at 5-10dB above
5. Labelling inconclusive responses as clear/definite and allowing them to influence the result

Improving quality

- Regular 1-day ABR “refresher” courses for all staff
 - Helps to introduce new protocols and allows staff to benchmark their interpretation skills
- Expert phone feedback to sites scoring poorly on QA
- On-site expert training when needed
- Mentoring struggling selected sites
 - all ABR cases reviewed via email until “signed off” as safe
- Introduction of local peer review networks

Local Peer Review Networks

There are a number of models but desirable features include:

- cases selected systematically (e.g. first 2 every month)
- reviewers receive training
- rotation of reviewers with sufficient cover for holiday and sickness
- a quick (typically 48-72 hours) review turnaround time for specific cases to allow feedback to be used in the management of that case
- use of a standard review format plus phone discussions between reviewer and tester when helpful
- regular meetings between testers & reviewers to check and benchmark their own practice
- access to experts for disagreements, queries and periodic moderation
- annual report to include an audit and review of arrangements and results of an external moderation

Local Peer Review Networks

- What doesn't work:
 - reviewing every case (too time consuming)
 - local tester selection of cases (usually only good ones chosen)
 - unwillingness to accept criticism of others
 - meeting over coffee every 3 months (ineffective)
 - reviewers not following national guidance
- The perfect compromise?
 - we don't yet know
 - several schemes are currently being piloted
 - watch is space!

Summary

- Quality standards and audit are essential
 - if our experience is typical you may have an unpleasant surprise!
- Stringent protocols deliver high quality
 - and are easier to achieve than most testers fear
- Staff training & support is vital
 - and should include additional support for struggling sites
- Peer review is likely to be good value
 - but we haven't yet identified the optimum model

Visit Us!

The English Newborn Hearing
Screening Programme

<http://hearing.screening.nhs.uk/>